

Proposed

Landon State Office Building
900 SW Jackson Street, Room 1031
Topeka, KS 66612-1228



phone: 785-296-7296
fax: 785-296-6212
www.ksbems.org

Dr. Joel E Hornung, Chair
Joseph House, Executive Director

Sam Brownback, Governor

**STATE OF KANSAS
BOARD OF EMERGENCY MEDICAL SERVICES
NOTICE OF HEARING ON AMENDED ADMINISTRATIVE REGULATION**

A public hearing will be conducted at 9:00 a.m. Tuesday, May 23, 2017 in Room 509 of the Landon State Office Building, 900 S.W. Jackson, Topeka, Kansas 66612 to consider the adoption of amended administrative regulation.

This 60-day notice of the public hearing shall constitute a public comment period for the purpose of receiving written public comments on the proposed amendments to the regulation. All interested parties may submit written comments prior to the hearing to the Board of Emergency Medical Services, Room 1031, Landon State Office Building, 900 S.W. Jackson, Topeka, Kansas 66612 or by email to james.reed@ks.gov. All interested parties will be given a reasonable opportunity to present their views orally on the adoption of the proposed amendments to the regulation during the hearing. In order to give all parties an opportunity to present their views, it may be necessary to request that each participant limit any oral presentations to five minutes.

Any individual with a disability may request accommodation in order to participate in the public hearing and may request the proposed regulation and economic impact statement in an accessible format. Requests for accommodation should be made at least five working days in advance of the hearing by contacting Ann Stevenson at 785-296-7296. Handicapped parking is located in front of and to the north of the Landon State Office Building.

RECEIVED

MAR 13 2017


KRIS W. KOBACH
SECRETARY OF STATE

~~Proposed~~

The regulation is being proposed for adoption on a permanent basis. A summary of the proposed regulation follows:

109-2-8. Standards for ground ambulances and equipment, is a current regulation that is being revised to only require equipment and supplies that are required by an ambulance service's scope of practice and protocols, and minor cleanup to language. K.A.R. 109-2-8(l), K.A.R. 109-2-8(m), K.A.R. 109-2-8(n), K.A.R. 109-2-8(o), K.A.R. 109-2-8(p).

There will be no anticipated economic impact to the Kansas Board of Emergency Medical Services associated with this regulatory revision. There will be no economic impact on other governmental agencies, private business or individuals with this regulatory revision.



109-2-8. Standards for ground ambulances and equipment. (a) Each ground ambulance shall meet the vehicle and equipment standards that are applicable to that type of ambulance.

(b) Each ground ambulance shall have the ambulance license prominently displayed in the patient compartment.

(c) The patient compartment size shall meet or exceed the following specifications:


(1) Headroom: 60 inches; and

(2) length: 116 inches.

(d) Each ambulance shall have a heating and cooling system that is controlled separately for the patient and the driver compartments. The air conditioners for each compartment shall have separate evaporators.

(e) Each ambulance shall have separate ventilation systems for the driver and patient compartments. These systems shall be separately controlled within each compartment. Fresh air intakes shall be located in the most practical, contaminant-free air space on the ambulance. The patient compartment shall be ventilated through the heating and cooling systems.

(f) The patient compartment in each ambulance shall have adequate lighting so that patient care can be given and the patient's status monitored without the need for portable or hand-held lighting. A reduced lighting level shall also be provided. A patient compartment light and step-well light shall be automatically activated by opening the entrance doors. Interior light fixtures shall be recessed and shall not protrude more than 1 1/2 inches.



APPROVED
FEB 28 2017
ATTORNEY GENERAL

APPROVED
FEB 22 2017
DEPT. OF ADMINISTRATION

(g) Each ambulance shall have an electrical system to meet maximum demand of the electrical specifications of the vehicle. All conversion equipment shall have individual fusing that is separate from the chassis fuse system.


(h) Each ground ambulance shall have lights and sirens as required by K.S.A. 8-1720 and K.S.A. 8-1738, and amendments thereto.

(i) Each ground ambulance shall have an exterior patient loading light over the rear door, which shall be activated both manually by an inside switch and automatically when the door is opened.

(j) The operator shall mark each ground ambulance licensed by the board as follows:

(1) The name of the ambulance service shall be in block letters, not less than four inches in height, and in a color that contrasts with the background color. The service name shall be located on both sides of the ambulance and shall be placed in such a manner that it is readily identifiable to other motor vehicle operators.

(2) Any operator may use a decal or logo that identifies the ambulance service in place of lettering. The decal or logo shall be at least 10 inches in height and shall be in a color that contrasts with the background color. The decal or logo shall be located on both sides of the ambulance and shall be placed in such a manner that the decal or logo is readily identifiable to other motor vehicle operators.



APPROVED
FEB 23 2017
ATTORNEY GENERAL

APPROVED
FEB 22 2017
DEPT. OF ADMINISTRATION

(3) Each ground ambulance initially licensed by the board before January 1, 1995 that is identified either by letters or a logo on both sides of the ground ambulance shall be exempt from the minimum size requirements in paragraphs (1) and (2) of this subsection.

(k) Each ground ambulance shall have a communications system that is readily accessible to both the attendant and the driver and is in compliance with K.A.R. 109-2-5(a).

(l) An operator shall equip each ground ambulance as follows:

(1) At least two annually inspected ABC fire extinguishers or comparable fire extinguishers with at least five pounds of dry chemical, which shall be secured. One fire extinguisher shall be easily accessible by the driver, and the other shall be easily accessible by the attendant;

(2) either two portable, functional flashlights or one flashlight and one spotlight;

(3) one four-wheeled or six-wheeled, all-purpose, multilevel cot with an elevating head and at least two safety straps with locking mechanisms;


(4) one urinal;

(5) one bedpan;

(6) one emesis basin or convenience bag;

(7) one complete change of linen;

(8) two blankets;



APPROVED
FEB 23 2017
ATTORNEY GENERAL

APPROVED
FEB 22 2017
DEPT. OF ADMINISTRATION

Approved

K.A.R. 109-2-8
Page 4


- (9) one waterproof cot cover;
- (10) one pillow; and
- (11) a no-smoking sign posted in the patient compartment and the driver compartment; and
- (12) mass-casualty triage tags.

(m) The operator shall equip each ground ambulance with the following internal medical systems:

(1) An oxygen system with at least two outlets located within the patient compartment and at least 2,000 liters of storage capacity, with a minimum oxygen level of 200 psi. The cylinder shall be in a compartment that is vented to the outside. The pressure gauge and regulator control valve shall be readily accessible to the attendant from inside the patient compartment; and

(2) a functioning, on-board, electrically powered suction aspirator system with a vacuum of at least 300 millimeters of mercury at the catheter tip. The unit shall be easily accessible with large-bore, nonkinking suction tubing and a large-bore, semirigid, ~~nonmetallic~~ nonmetallic oropharyngeal suction tip.

(n) The operator shall equip each ground ambulance with the following medical equipment:



APPROVED
FEB 23 2017
ATTORNEY GENERAL

APPROVED
FEB 22 2017
DEPT. OF ADMINISTRATION

(1) A portable oxygen unit of at least 300-liter storage capacity, complete with pressure gauge and flowmeter and with a minimum oxygen level of 200 psi. The unit shall be readily accessible from inside the patient compartment;

(2) a functioning, portable, self-contained battery or manual suction aspirator with a vacuum of at least 300 millimeters of mercury at the catheter tip and a transparent or translucent collection bottle or bag. The unit shall be fitted with large-bore, nonkinking suction tubing and a large-bore, semirigid, nonmetallic oropharyngeal suction tip, unless the unit is self-contained; and

~~(3) a hand-operated, adult bag-mask ventilation unit, which shall be capable of use with the oxygen supply;~~

~~(4) a hand-operated, pediatric bag-mask ventilation unit, which shall be capable of use with oxygen supply;~~


~~(5) oxygen masks in adult and pediatric sizes;~~

~~(6) nasal cannulas in adult and pediatric sizes;~~

~~(7) oropharyngeal airways in adult, pediatric, and infant sizes;~~

~~(8) a blood pressure manometer with extra-large, adult, and pediatric cuffs and a stethoscope;~~

~~(9) an obstetric kit with contents as described in the ambulance service's medical protocol;~~



APPROVED
FEB 23 2017
ATTORNEY GENERAL

APPROVED
FEB 22 2017
DEPT. OF ADMINISTRATION

~~(10) sterile burn sheets;~~
~~(11) sterile large trauma dressings;~~
~~(12) assorted sterile gauze pads;~~
~~(13) occlusive gauze pads;~~
~~(14) rolled, self-adhering bandages;~~
~~(15) adhesive tape at least one inch wide;~~
~~(16) bandage shears;~~
~~(17) one liter of sterile water, currently dated or one liter of sterile saline, currently dated; and~~

~~(18) currently dated supplies, medications, and equipment as authorized by the scope of practice and protocols, in accordance with the applicable list of supplies, medications, and equipment approved by the medical director.~~

~~(o) The operator shall equip each ground ambulance with the following patient handling and splinting equipment:~~

~~(1) If required by protocol, a long spinal immobilization device, complete with accessories to immobilize a patient;~~

~~(2) a set of extremity splints including one arm and one leg splint, in adult and pediatric sizes;~~

~~(3) a set of rigid cervical collars in assorted adult and pediatric sizes;~~

~~(4) If required by protocol, foam wedges or other devices that serve to stabilize the head, neck, and back as one unit; and~~



APPROVED
FEB 23 2017
ATTORNEY GENERAL

APPROVED
FEB 22 2017
DEPT. OF ADMINISTRATION

~~(5)~~ patient disaster tags;

~~(p)~~ The operator shall equip each ground ambulance with the following blood-borne and body fluid pathogen protection equipment in a quantity sufficient for crew members:

(1) Surgical or medical protective gloves;

(2) protective goggles, glasses or chin-length clear face shields;


(3) filtering masks that cover the mouth and nose;

(4) nonpermeable, full-length, long-sleeve protective gowns;

(5) a leakproof, rigid container clearly marked as "Biohazard" for the disposal of sharp objects; and

(6) a leakproof, closeable container for soiled linen and supplies.

~~(q)~~ (p) If an operator's medical protocols or equipment list is amended, a copy of these changes shall be submitted to the board by the ambulance service operator within 15 days of implementation of the change. Equipment and supplies obtained on a trial basis or for temporary use by the operator shall not be required to be reported to the board by an operator. (Authorized by K.S.A. ~~2015~~ 2016 Supp. 65-6110; implementing K.S.A. ~~2015~~ 2016 Supp. 65-6110 and K.S.A. 65-6128; effective May 1, 1985; amended, T-88-24, July 15, 1987; amended May 1, 1988; amended July 17, 1989; amended Aug. 16, 1993; amended Jan. 31, 1997; amended Jan. 27, 2012; amended Feb. 13, 2015; amended April 29, 2016; amended P-_____.)



APPROVED
FEB 23 2017
ATTORNEY GENERAL

APPROVED
FEB 22 2017
DEPT. OF ADMINISTRATION

Landon State Office Building
900 SW Jackson Street, Room 1031
Topeka, KS 66612-1228



Proposed

phone: 785-296-7296
fax: 785-296-6212
www.ksbems.org

Dr. Joel E. Hornung, Chair
Joseph House, Executive Director

Sam Brownback, Governor

ECONOMIC IMPACT STATEMENT K.A.R. 109-2-8

I. Summary of Proposed Regulation, Including Its Purpose.


K.A.R. 109-2-8 is a current regulation that defines requirements for supplies and equipment on a ground ambulance. This regulation is being revised to remove some supplies and equipment from the required list, and only require supplies and equipment as required by an ambulance service's scope of practice and medical protocols; K.A.R. 109-2-8(l), K.A.R. 109-2-8(m), K.A.R. 109-2-8(n), K.A.R. 109-2-8(o), K.A.R. 109-2-8(p).

II. Reason or Reasons the Proposed Regulation Is Required, Including Whether or Not the Regulation Is Mandated by Federal Law.

This regulation is being revised so ambulance services will only be required to carry supplies required by an ambulance service's scope of practice and medical protocols. This revision removes requirements for an ambulance service to carry equipment and supplies for which they have no medical protocols. This regulation is not mandated by federal law.

III. Anticipated Economic Impact upon the Kansas Board of Emergency Medical Services.

There will be no anticipated economic impact to the Kansas Board of Emergency Medical Services associated with this regulatory revision.



IV. Anticipated Financial Impact upon Other Governmental Agencies and upon Private Business or Individuals.

There will be no anticipated economic impact on other governmental agencies, private business or individuals with this regulatory revision.

V. Anticipated Economic Impact upon Consumers of the Services Subject to the Regulation or Its Enforcement.

There will be no anticipated economic impact on Emergency Medical Services attendants, agencies or other Emergency Medical Services related entities associated with this regulatory revision.

VI. Less Costly or Intrusive Methods That Were Considered, but Rejected, and the Reason for Rejection.

There are no anticipated costs or intrusive methods associated with the revision of this regulation.

